

Local physical chemistry statements for low cost sensor added value

Scheffler, Janice; Vieno, Massimo; Flynn, Michael; Lacy, Stuart; Diez, Sebastian; Pope, Francis D.; Edwards, Pete; Braban, Christine F.

DOI:

[10.5194/egusphere-egu22-12342](https://doi.org/10.5194/egusphere-egu22-12342)

License:

Creative Commons: Attribution (CC BY)

Document Version

Publisher's PDF, also known as Version of record

Citation for published version (Harvard):

Scheffler, J, Vieno, M, Flynn, M, Lacy, S, Diez, S, Pope, FD, Edwards, P & Braban, CF 2022, Local physical chemistry statements for low cost sensor added value. in *EGU General Assembly 2022.*, EGU22-12342, European Geosciences Union, EGU General Assembly 2022, Vienna, Austria, 23/05/22.
<https://doi.org/10.5194/egusphere-egu22-12342>

[Link to publication on Research at Birmingham portal](#)

General rights

Unless a licence is specified above, all rights (including copyright and moral rights) in this document are retained by the authors and/or the copyright holders. The express permission of the copyright holder must be obtained for any use of this material other than for purposes permitted by law.

- Users may freely distribute the URL that is used to identify this publication.
- Users may download and/or print one copy of the publication from the University of Birmingham research portal for the purpose of private study or non-commercial research.
- User may use extracts from the document in line with the concept of 'fair dealing' under the Copyright, Designs and Patents Act 1988 (?)
- Users may not further distribute the material nor use it for the purposes of commercial gain.

Where a licence is displayed above, please note the terms and conditions of the licence govern your use of this document.

When citing, please reference the published version.

Take down policy

While the University of Birmingham exercises care and attention in making items available there are rare occasions when an item has been uploaded in error or has been deemed to be commercially or otherwise sensitive.

If you believe that this is the case for this document, please contact UBIRA@lists.bham.ac.uk providing details and we will remove access to the work immediately and investigate.



Local Physical Chemistry Statements for Low cost sensor added value

Janice Scheffler¹, Massimo Vieno¹, Michael Flynn², Stuart Lacy³, Sebastian Diez³, Francis D. Pope⁴, Pete Edwards³, and Christine F. Braban¹

¹UK Centre for Ecology and Hydrology, Penicuik, United Kingdom

²University of Manchester, Manchester, United Kingdom

³University of York, York, United Kingdom

⁴University of Birmingham, Birmingham, United Kingdom

Air quality monitoring with a high spatial and temporal resolution is essential to understand the sources, processes, and impacts of air quality (AQ) on human health and the environment, especially in densely populated urban areas. Current state-of-the-art instruments don't allow for such a high spatial resolution AQ monitoring due to costs. Low cost sensors (LCS) provide an opportunity to bridge this divide. There is a large volume of research papers assessing the performance of LCS for particulate matter (PM) against reference PM instrumentation. Across most studies, a general observation is that LCS PM sensors as offered by the commercial and research markets have an inherent problem with accuracy at high relative humidities and for varying PM composition.

This study uses measured and modelled chemical composition of PM_{2.5} at the Manchester UK urban supersite to address the underlying physical chemistry which leads to the mass of the measured PM. Measurements from LCS and reference PM instruments co-located with state-of-the-art research chemical composition and meteorology. To calculate the composition of PM_{2.5}, we used XACT for metals, the ACSM for salt ion concentrations, and an Aethalometer for Black Carbon. From this, the LCSs variability in data can be understood. We present a comparison of LCS PM_{2.5} concentrations with measured and calculated total PM_{2.5} concentrations from reference instruments for time periods with different Air Quality characteristics and discuss the physico-chemical characteristics leading to varying results of low cost PM sensors. The future research will apply methodology to modelled UK wide data and develop tools for non specialist users to understand LCS in terms of local air pollution and specifically PM_{2.5} chemical composition and meteorology.