

Editorial

Helgason, Benedikt; Bryant, Michael G.; Ferguson, Stephen J.; Hall, Richard M.

DOI:

[10.3389/fbioe.2023.1264118](https://doi.org/10.3389/fbioe.2023.1264118)

License:

Creative Commons: Attribution (CC BY)

Document Version

Publisher's PDF, also known as Version of record

Citation for published version (Harvard):

Helgason, B, Bryant, MG, Ferguson, SJ & Hall, RM 2023, 'Editorial: Advanced pre-clinical and pre-surgical assessment of musculo-skeletal medical devices', *Frontiers in Bioengineering and Biotechnology*, vol. 11, 1264118. <https://doi.org/10.3389/fbioe.2023.1264118>

[Link to publication on Research at Birmingham portal](#)

General rights

Unless a licence is specified above, all rights (including copyright and moral rights) in this document are retained by the authors and/or the copyright holders. The express permission of the copyright holder must be obtained for any use of this material other than for purposes permitted by law.

- Users may freely distribute the URL that is used to identify this publication.
- Users may download and/or print one copy of the publication from the University of Birmingham research portal for the purpose of private study or non-commercial research.
- User may use extracts from the document in line with the concept of 'fair dealing' under the Copyright, Designs and Patents Act 1988 (?)
- Users may not further distribute the material nor use it for the purposes of commercial gain.

Where a licence is displayed above, please note the terms and conditions of the licence govern your use of this document.

When citing, please reference the published version.

Take down policy

While the University of Birmingham exercises care and attention in making items available there are rare occasions when an item has been uploaded in error or has been deemed to be commercially or otherwise sensitive.

If you believe that this is the case for this document, please contact UBIRA@lists.bham.ac.uk providing details and we will remove access to the work immediately and investigate.



OPEN ACCESS

EDITED AND REVIEWED BY
Markus O. Heller,
University of Southampton,
United Kingdom

*CORRESPONDENCE
Benedikt Helgason,
✉ bhelgason@ethz.ch

RECEIVED 20 July 2023
ACCEPTED 26 July 2023
PUBLISHED 21 July 2023

CITATION

Helgason B, Bryant MG, Ferguson SJ and Hall RM (2023), Editorial: Advanced pre-clinical and pre-surgical assessment of musculo-skeletal medical devices. *Front. Bioeng. Biotechnol.* 11:1264118. doi: 10.3389/fbioe.2023.1264118

COPYRIGHT

© 2023 Helgason, Bryant, Ferguson and Hall. This is an open-access article distributed under the terms of the [Creative Commons Attribution License \(CC BY\)](https://creativecommons.org/licenses/by/4.0/). The use, distribution or reproduction in other forums is permitted, provided the original author(s) and the copyright owner(s) are credited and that the original publication in this journal is cited, in accordance with accepted academic practice. No use, distribution or reproduction is permitted which does not comply with these terms.

Editorial: Advanced pre-clinical and pre-surgical assessment of musculo-skeletal medical devices

Benedikt Helgason^{1,2*}, Michael G. Bryant³, Stephen J. Ferguson^{1,2} and Richard M. Hall³

¹Institute for Biomechanics, ETH Zurich, Zurich, Switzerland, ²Future Health Technologies, Singapore-ETH Centre, Campus for Research Excellence and Technological Enterprise (CREATE), Singapore, Singapore, ³School of Mechanical Engineering, Faculty of Engineering and Physical Sciences, University of Leeds, Leeds, United Kingdom

KEYWORDS

medical devices, pre-clinical, pre-surgical, testing, *in silico*, *ex vivo*, *in vivo*

Editorial on the Research Topic

Advanced pre-clinical and pre-surgical assessment of musculo-skeletal medical devices

The orthopaedic implant market is worth in excess of €30B per annum globally with articulating joint replacements representing the largest orthopaedic sector. Growth in the sector of between 2% and 5% is expected over the near term due to the rapid demographic shift. Pre-clinical assessment, both experimental and *in silico*, is a series of necessary steps for the development, optimisation and validation of medical devices. It comprises the testing of implants using assessments that conform to agreed standards as well as those that are more bespoke and focus on the specific requirements and perceived usage of the device. However, once introduced the new technologies are not always successful clinically and may harm the patient. This may require an expensive intervention to correct the loss in the patient's quality of life and greater mortality risk. These deficits in implant outcomes have brought into focus the role of pre-clinical simulation and the wider regulatory science that supports these activities.

The theme of the research included in this Research Topic, which is driven by our experience within the EU MSCA European Training Network, BioTrib, is how best to facilitate improvements in pre-clinical *in silico*, *in vitro* and *in vivo* testing within an orthopaedic context to allow a reduction in the adverse events that occur in implants once marketed. The studies cover techniques such as artificial intelligence, image processing, laboratory and computer simulations, to achieve this end.

Several of the studies are focused on addressing the limitations of using a "human in the loop" for, e.g., diagnosis or implant selection. [Qu et al.](#) used a deep learning approach for localizing cruciate ligament rupture on knee MRI images. The overall aim of such work is to identify the location and tissue quality of injured ACLs to determine if the ACL repair surgery can be performed. The goal of the work of [Burge et al.](#) was to develop a computational tool for automatic selection of total knee replacement implant size using X-ray images. Both studies support the use of algorithms to achieve repeatable outcomes to reduce operator dependency.

Some of the studies are comparing the biomechanical efficacy of two or more surgical procedures or devices, with the overall aim of ranking the procedures or improving device design. Here the benefit of preclinical *in silico* testing can be utilized to avoid subjecting patients to clinical trials of devices or procedures that may involve elevated risk. [Fung et al.](#), e.g., studied the efficacy

of prophylactic ceramic-based cement augmentation of the proximal femur to prevent hip fracture. The authors provide insight into the potential efficacy of such a procedure, but prophylaxis for fracture prevention is a controversial Research Topic among clinicians. Using an advanced experimental model, [Techens et al.](#) studied the biomechanical consequences of cement discoplasty. This surgical procedure is relatively novel, and experimental models that can provide insight into, e.g., the changes in spine kinematics following the treatment, are needed.

Another important aspect of pre-clinical testing of orthopaedic devices is addressed in the study by [Kohli et al.](#) Here, a bioreactor analyses of tissue ingrowth, ongrowth and remodelling around implants is used to reduce the need for animal testing. Some of the studies in this Research Topic, fall under the category of novel measurement techniques and protocols. In the study of [Kandel et al.](#), an automated system is introduced for polymer wear debris analysis in total disc arthroplasty using convolution neural network. Using dual-plane fluoroscopy to observe treadmill walking, [Zhou et al.](#) investigate whether multi-planar instability, laxity and reduced knee flexion during the support phase of walking are determinants of return to sports. On the other hand, [D'Isidoro et al.](#) use dual plane moving fluoroscopy-based analysis, to study total hip arthroplasty kinematics during unrestricted activities of daily living. These advanced measurement techniques are providing novel insights into aspects of total joint replacements that are difficult to achieve with other, more conventional methods.

Overall, the papers on this Research Topic cover some recent advances in preclinical testing of musculoskeletal devices. The editors hope that this Research Topic will contribute to advancing the field of orthopaedics, by inspiring researchers, clinicians and industry alike, towards reduction in the need for animal testing, improving patient safety and lowering healthcare costs.

Author contributions

BH: Writing–review and editing. MB: Writing–review and editing. SF: Writing–review and editing. RH: Writing–review and editing.

Funding

This project has received funding from the European Union's Horizon 2020 research and innovation programme under the Marie Skłodowska-Curie grant agreement No 956004.

Conflict of interest

The authors declare that the research was conducted in the absence of any commercial or financial relationships that could be construed as a potential conflict of interest.

The author(s) declared that they were an editorial board member of *Frontiers*, at the time of submission. This had no impact on the peer review process and the final decision

Publisher's note

All claims expressed in this article are solely those of the authors and do not necessarily represent those of their affiliated organizations, or those of the publisher, the editors and the reviewers. Any product that may be evaluated in this article, or claim that may be made by its manufacturer, is not guaranteed or endorsed by the publisher.